

INDIANA DEPARTMENT OF TRANSPORTATION DIVISION OF MATERIALS AND TESTS

VERIFYING CALIBRATION SETTINGS FOR SUPERPAVE GYRATORY COMPACTORS ITM No. 908-22

1.0 SCOPE.

- 1.1 This test method covers the procedures for verifying calibration settings on an approved SUPERPAVE Gyratory Compactor (SGC).
- 1.2 This ITM may involve hazardous materials, operations, equipment and may not address all of the safety problems associated with the use of the test method. The user of the ITM is responsible for establishing appropriate safety and health practices and to determine the applicability of regulatory limitations prior to use.
- **TERMINOLOGY.** Definitions for terms and abbreviations shall be in accordance with the Department's Standard Specifications, Section 101.
- **3.0 SIGNIFICANCE AND USE.** This ITM is used by laboratory personnel to verify the calibration settings on an approved SGC.

4.0 APPARATUS.

- **4.1** Digital Stopwatch, readable to 1 s
- 4.2 Dynamometer or load cell, National Institute of Standards and Technology (NIST) traceable in accordance with ASTM E74 and calibrated every 24 months
- **4.3** Precision gauge blocks, NIST traceable and calibrated every 24 months
- **4.4** Device for verifying the external angle of gyration

5.0 PROCEDURES.

5.1 General.

- **5.1.1** Use the manufacturer's calibration equipment and procedures for verification of the speed of gyration, ram pressure, ram position, and external angle of gyration.
 - AFG2AS gyratory devices manufactured by Pine Test Equipment, Inc. will have the external angle of gyration verified by the manufacturer as part of the annual verification process.
- **5.1.2** Turn on the SGC and allow the SGC to warm up for approximately 15 minutes or the manufacturer's recommended time prior to verifying the calibration settings.

ITM 908-22 Revised 11/03/2022

5.1.3 If any of the calibration settings are outside the specified tolerances of 6.0, a new calibration shall be performed for all of the parameters.

5.2 Speed of Gyration.

- **5.2.1** Set the SGC for 30 revolutions and start the machine
- **5.2.2** Record the time for the 30 gyrations using the stopwatch

5.3 Ram Pressure.

- **5.3.1** Center the dynamometer or load cell under the ram
- **5.3.2** If the dynamometer is used, flex the device to the maximum load required during calibration
- **5.3.3** Apply a force with the SGC. The SGC shall load the ram to a minimum of two forces spanning the full capacity of the SGC.
- **5.3.4** Record the force indicated on the load cell. If the dynamometer is used, record the deflection and determine the force from the conversion chart of the dynamometer.

5.4 Ram Position.

- **5.4.1** Insert the gauge blocks under the ram so that the height of the blocks is 6 in.
- **5.4.2** Apply a force with the SGC
- **5.4.3** Record the height indicated on the control panel

5.5 Angle of Gyration.

- **5.5.1** Attach the appropriate indicator for determination of the external angle of gyration
- **5.5.2** Verify that the measurement indicates that the allowable tolerance of the external angle is within $\pm 0.02^{\circ}$

6.0 TOLERANCES.

6.1 The speed of gyration shall be 30 gyrations in 60 ± 1 s.

ITM 908-22 Revised 11/03/2022

6.2 The ram pressure shall be within \pm 3 percent of the pressure indicated on the SGC.

- 6.3 The ram position shall be within \pm 0.004 in. of the height of the gauge blocks.
- 6.4 The external angle shall be $\pm 0.02^{\circ}$ from the established external angle. The dial indicator value shall be within the verification range.

<< The remainder of this page is intentionally blank >>

SUPERPAVE Gyratory Compactor Verification ITM 908

SGC Model:	Ser	rial No: _	Machine Hours:					
Dynamometer/Load Cell Serial No.:					Calibration Date:			
Gauge Block Serial Nos.:				Calibration Date:				
Speed of Gyration	Ram Pressure							
30 gyrations	Applied Load – Load Cell			Applied Load - Dynamometer				
$in 60 \pm 1s$	Displayed	Actual		Deflection	Force		Actual	
Ram Position				External Angle				
Height			Dial Indicator Value Verification					
Displayed	d Actual			(Second - First = Value)			Range	
			Second: First:					
	Value: Second:							
	First:			st:	_			
			Value:					
Remarks:								
Verified by:								
Date:		Next Due Date:						